



Low Organic Water

PRODUCT NUMBER: S010901

LOT NUMBER: 9108080

RELEASE DATE: August, 2008

EXPIRY DATE: August, 2011

CERTIFICATE OF ANALYSIS

<u>Tests</u>	<u>Units</u>	<u>Value</u>
Colour	APHA	< 5
Residue after Evaporation	ppm	< 1
Fluorescence (as Quinine)	ppt	< 100
Resistivity (at time of bottling)	MΩ	> 18
Total Organic Carbon	ppb	< 20

LC Gradient Suitability Test:	passes test
Column: 25 cm C18	
Flow rate: 2 mL/min	
Conditions: 40 ml of water are passed through the column and eluted with a water acetonitrile gradient at 5% per min.	
Max. Abs.: 0.002 at 254 nm	

Optical Absorbance	Wavelength	Absorbance
	200 nm	< 0.01
	254 nm	< 0.005
	300 nm	< 0.005
	400 nm	< 0.005


Dr. B. McKelvey
QA/QC Manager

 **SEASTAR CHEMICALS INC**
A member of the AXYS Group

10005 McDonald Park Road, Sidney, BC Canada V8L 5Y2
Phone: 250-655-5880 - Fax: 250-655-5888 - Toll Free: 1-800-663-2330
www.seastarchemicals.com